

---

STATUTORY INSTRUMENTS

---

**1986 No. 1685**

The Weights and Measures (Local and Working Standard  
Weights and Testing Equipment) Regulations 1986

*Prescribed limits of error*

- 8.**—(1) The prescribed limits of error relating to local and working standard weights and test weights shall be those set out in Tables 1 to 6 in Schedule 2 to these Regulations.
- (2) The prescribed limit of error relating to egg grading machine test poises shall be 35 mg.